

Department of Forensic Science
Breath Alcohol Section
Instrument Location History
1/1/1990 To 12/3/2014

KMS
gnd

Instrument Serial Number: 010461

As of 03-Dec-14

Start Date	End Date	Agency Name
03-Dec-14		DFS Central Lab
29-Aug-14	03-Dec-14	Franklin County SO
20-Aug-14	29-Aug-14	DFS Central Lab
18-Jun-14	20-Aug-14	Franklin County SO
29-May-14	18-Jun-14	DFS Central Lab
26-Jul-12	29-May-14	Appomattox County SO
10-Jul-12	26-Jul-12	DFS Central Lab
09-Dec-11	10-Jul-12	Henrico County PD
09-Dec-11	09-Dec-11	DFS Central Lab
17-Aug-10	09-Dec-11	Henrico County PD
25-Mar-09	17-Aug-10	DFS Central Lab
28-Oct-08	25-Mar-09	Russell County SO
08-May-08	28-Oct-08	DFS Central Lab

Handwritten initials/signature in the top right corner.

**INTOX EC/IR II
Quality Assurance Worksheet**

Instrument Serial Number **010461** Worksheet Start Date **12/3/2014**
Location **Franklin County**
Address **40 E. Court St., Rocky Mount, VA 24151**
DFS Technician **Heather Stanton** License No. **18910**
☐ Laboratory ☒ On-Site

Site Specification: No detrimental environmental conditions exist. ☒

Instrument Barometer (mm HG) **735** Reference Barometer (mm HG) **735**
Reference Barometer(RB)Serial # **009113** RB Calibration Due **4/28/2015**

Measurement Assurance Check

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.294
0.300	0.290	0.281	0.298	Sample 2	0.294
Precision		sample min	sample max	Sample 3	0.293
0.001		0.293	0.294		

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.097
0.100	0.097	0.094	0.100	Sample 2	0.097
Precision		sample min	sample max	Sample 3	0.097
0		0.097	0.097		

Dry gas standard Lot No. (with tank no.) **AG332301-27**

- ☐ Replaced dry gas standard (+O-ring)
- ☐ Installed at Location
- ☒ Removed to DFS-Central

Supplies
Mouthpieces
Certificates of Analysis
Operator Worksheet
Other:



Notes:

[Large blank area for handwritten notes]

Instrument Serial Number

010461

Certification Date

☐ Calibrated☐ Certified☒ Measurement Assurance Check☒ Instrument Test☐ Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician

Heath W. Stach

Date


12/3/14

Issuing Analyst

JPD
JPD

Date

12/4/14




COMMONWEALTH OF VIRGINIA
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

the
Fed

NAME OF ACCUSED INSTRUMENT, TEST,	NAME OF COURT DFS
---	-----------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY STANTON, HEATHER, M		AGENCY DFS Central Lab
DFS LICENSE NUMBER 18910	LICENSE EXPIRES 10/01/2016	DATE TEST CONDUCTED 12/03/2014
TEST EQUIPMENT NUMBER 010461		

RESULTS: TIME SAMPLE TAKEN **10:50 EST**

SAMPLE'S ALCOHOL CONTENT **0.00** GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS _____ DAY OF _____, 20 _____.

BREATH TEST OPERATOR

☐ I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS _____ SUBJECT'S SIGNATURE

☐ SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS _____ OPERATOR'S SIGNATURE



IntoxNet MIS Report

Report Generated 03 Dec 2014 at 11:19

Ad
AMS

Test Results

Instrument Serial Number 010461

Test # 002468

Subject Test

Test Location 1 Department of
Test Date 03 Dec 2014

Test Location 2 Forensic Science
Test Time 10:44

Test Location 3
System Check Passed

Operator's Last Name STANTON
Agency DFS Central Lab

Operator's First Name HEATHER

Operator's Middle Initial M

Card Serial Number 118910

License Number 18910

Effective Date 10/01/2014

Expiration Date 10/01/2016

Subject's Last Name INSTRUMENT

Subject's First Name TEST

Subject's Middle Initial

Subject's Date of Birth 00/00/0000

Subject's Sex Male

Driver's License Number

Driver's License Expiration 00/00/0000

Driver's License State

Court Name DFS

End Date 03 Dec 2014 End Time 10:51

Result Time 10:50

Result Date 03 Dec 2014 Result 0.00

Data Type DIAG

Sample Value Pass

Sample Time 10:44

Data Type BLK

Sample Value 0.000

Sample Time 10:45

Data Type CHK

Sample Value 0.097

Sample Time 10:45

Data Type BLK

Sample Value 0.000

Sample Time 10:47

Data Type SUBJ

Sample Value 0.000

Sample Time 10:48

Data Type BLK

Sample Value 0.000

Sample Time 10:49

Data Type SUBJ

Sample Value 0.000

Sample Time 10:50

Data Type BLK

Sample Value 0.000

Sample Time 10:51

Standard Type Dry Gas Std

Standard Value 0.097

Standard Lot Number AG332301-27

Standard Expiration Date 11/19/2015

Tank Pressure 345

Barometric Pressure 735 mmHg

Blow Sample Number 1 Blow Duration 3.33 sec

Blow Volume 1759 cc End-of-Blow Time 10:48

Blow Sample Number 2 Blow Duration 3.60 sec

Blow Volume 1833 cc End-of-Blow Time 10:50

Tamper Evident Stamp cac72584

Test Status Code 0

Test Status Success

Intox
Sed

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010461 Test Number: 2469

Test Date: 12/03/2014 Test Time: 10:52 EST

Dry Gas Target: 0.097

Lot Number: AG332301-27 Exp Date: 11/19/2015

Tank Pressure: 345 psi Barometric Pressure: 734 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	10:53
CHK	0.097	10:54
BLK	0.000	10:55
CHK	0.097	10:56
BLK	0.000	10:57
CHK	0.097	10:58

Test Status: *Success*

Calibration CRC: 69B6385F

Handwritten initials/signature in the top right corner.

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010461 Test Number: 2470

Test Date: 12/03/2014 Test Time: 11:04 EST

Dry Gas Target: 0.290

Lot Number: AG428003-11 Exp Date: 10/07/2016

Tank Pressure: 404 psi Barometric Pressure: 734 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	11:05
CHK	0.294	11:05
BLK	0.000	11:07
CHK	0.294	11:08
BLK	0.000	11:10
CHK	0.293	11:10

Test Status: *Success*

Calibration CRC: 69B6385F